

Uniphos Detector Tube Data Sheet

Direct Reading Length of Stain Type

Carbon Dioxide

CO₂

SCD-5.1

TWA (TLV): 5000 ppm

STEL (TLV): 30,000 ppm

Non Flammable

Specification:

Measurement Range	Extended	Standard	Extended
	0.015-0.25 %	0.03-0.5 %	0.06-1 %
No. of Pump Strokes	4 (400 mL)	2 (200 mL)	1 (100 mL)
Volume Correction Factor (VCF)*	0.5	1.0	2.0
Sampling Time	3.5 minutes per pump stroke (100 mL)		
Color Change	Blue → Off white		
Shelf Life	2 years		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	CO ₂ + NaOH → Na ₂ CO ₃ + H ₂ O		
Relative standard deviation	± 10 %		

* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
CO	3000	No	No effect
NH ₃	50000	No	No effect
NO ₂	200	No	No effect
SO ₂	2050	+	
n-Hexane	1500	No	No effect
Toluene	400	No	No effect

Correction for Environmental Parameters:

Temperature (°C) :	4	20	30	40
Temperature Correction Factor (TCF) *	0.9	1	1.1	1.2

Relative humidity	Not necessary between 10 - 90 %.
-------------------	----------------------------------

* Multiply the observed reading by the correction factor (VCFXTCF) to obtain the true concentration.



Manufactured By: Uniphos Envirotronic Pvt. Ltd.(100% EOU)
Nahuli (PO), Umbergaon (Tal.)
Valsad (Dist), Gujarat, India – 396108

Rev.1016