

# Uniphos Detector Tube Data Sheet

Direct Reading Length of Stain Type

## Nitric Acid

TWA (TLV): 2 ppm



Ceiling (TLV): 4 ppm

## SNA-2M

Flammable Range: NA %

### Specification:

Measurement Range	Extended	Standard	Extended
	0.5 – 10 ppm	1 – 20 ppm	2 – 40 ppm
No. of Pump Strokes	2 (200 mL)	1 (100 mL)	0.5 (50 mL)
Volume Correction Factor (VCF)*	0.5	1.0	2.0
Sampling Time	1 minute per pump stroke (100 mL)		
Color Change	Greenish yellow → Purple		
Shelf Life	1 year		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 10 mg/l		
Reaction Principle	HNO <sub>3</sub> + Base → Reaction product		
Relative standard deviation	± 10 %		

\* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

### Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
HCl		+	Greenish yellow to Purple
Cl <sub>2</sub>		+	Greenish yellow to Purple
CO	10	No	No Effect
NO <sub>2</sub>	10	No	No Effect
Methanol	100	No	No Effect

### Other Measurable Compounds:

Compound	Gas Correction factor (GCF)*	No. of pump strokes	Measuring range
Hydrogen Bromide	1.0	1	1-20 ppm

\* Multiply the observed reading by the correction factor (VCF x GCF) to obtain the true concentration.

### Correction for Environmental Parameters:

**Temperature** - Not necessary between 0 - 40°C (32 - 104°F).

Relative Humidity (mg/l):	6	8	10	12	16
Correction Factor (HCF)*:	0.53	0.72	1.0	1.2	1.41

\* Multiply the observed reading by the correction factors (VCFxHCF) to obtain the true concentration.



**Manufactured By:** Uniphos Envirotronic Pvt. Ltd.  
Nahuli (PO), Umbergaon (Tal.)  
Valsad (Dist), Gujarat, India – 396108

Rev.0717