

# Uniphos Detector Tube Data Sheet

Direct Reading Length of Stain Type

## Water Vapor

## H<sub>2</sub>O

## SWA-7M/lb

TWA (TLV): NA

STEL (TLV): NA

Non Flammable

### Specification:

Measurement Range	Extended	Standard	Extended
	Do not extend	62.5 – 1125 LB/MMCF	Do not extend
No. of Pump Strokes	1 (100 mL)		
Volume Correction Factor (VCF)*	1.0		
Sampling Time	1 minute per pump stroke (100 mL)		
Color Change	Yellow → Blue		
Shelf Life	1 year		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F).		
Reaction Principle	H <sub>2</sub> O + Mg (ClO <sub>4</sub> ) <sub>2</sub> → Mg(ClO <sub>4</sub> ) <sub>2</sub> .H <sub>2</sub> O		
Relative standard deviation	± 10 %		

\* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

### Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
CH <sub>4</sub>	99.9%	No	No effect
H <sub>2</sub> S	2000	No	No effect
Methanol	50	No	No effect
Hexane	3000	No	No effect

### Correction for Environmental Parameters:

Temperature	Not necessary between 0 - 40°C (32 - 104°F).
-------------	--

**Manufactured By:** Uniphos Envirotronic Pvt. Ltd.(100% EOU)  
Nahuli (PO), Umbergaon (Tal.)  
Valsad (Dist), Gujarat, India – 396108



Rev.0817